

Quality in Control


Breast Analyte Control^{DR}

Product Codes: HCL016, HCL017 and HCL018

Contents

- 1. Breast Analyte Control^{DR}
- 2. Breast Analyte Control^{DR} images: Her2
- 3. Her2 1+ versus 2+ cells
- 4. Breast Analyte Control^{DR} images: ER
- 5. Breast Analyte Control^{DR} images: PR
- 6. QC Scoring: ER/PR
- 7. Also available from HistoCyte Laboratories Ltd


HistoCyte Laboratories Ltd is based in the heart of the Newcastle University campus. Started in 2014 by scientists with a combined experience of over 30 years in the development of reagents for immunohistochemistry and in-situ hybridization. Collaborating with pathologists locally and globally, HistoCyte Laboratories Ltd is developing a range of cost effective products designed to help scientists to maintain and develop the quality of assays within their laboratory.




Breast Analyte Control^{DR}


Breast Analyte Control^{DR} is part of the **Dynamic Range** of HistoCyte Products. When a Dynamic Range or a control of high sensitivity is required the **Breast Analyte Control^{DR}** is ideal. This product contains four cells of varying expression for a variety of breast biomarkers, including Her2, Estrogen Receptor and Progesterone Receptor. It also includes an osteosarcoma negative control.


Breast Analyte Control^{DR} is available as pre-cut slides (2 slide and 5 slide mailers) and cell microarray blocks.





Breast Analyte Control^{DR} Her2





Her2 1+ versus 2+ Cells

Her2 2+ is difficult to obtain in laboratories and not always consistent. Tumour is often highly heterogeneous either because of the way the protein is expressed or because of the affects of fixation and processing.

2+ creates the greatest degree of inter-observer discordance¹.



"Is it a weak 2+ or a strong 1+?"

There is no linear correlation between 0/1+/2+ and 3+ (orange vs blue lines).

The linear relationship between 0, 1+ and 2+, makes 1+ cell line is the best predictor of assay drift (red line).

There are two 1+ cells in the Breast Analyte Control^{DR} allowing a laboratory to demonstrate consistency within the control.


Diagnostic algorithms are not necessarily a means to assess Quality. While Her2 are scored 0, 1+, 2+, 3+ controls in laboratories are often scored using an adapted system to convey how the assay has performed. See table below.


	QC score						
Her2	0	0/1+	1+	1/2+	2+	2/3+	3+

This "grey" area is necessary to impart to the pathologist how the laboratory feel the assay has performed. This QC score can be reflected on by the pathologist scoring the case. This is where the score has to be within the 0, 1+, 2+, 3+. There can be no grey area.


^{1.} Turashvili G, Leung S, Turbin D, Montgomery K, Gilks B, West R, Carrier M, Huntsman D, Aparicio S. Inter-observer reproducibility of Her 2 immunohistochemical assessment and concordance with fluorescent in situ hybridization (FISH): pathologist assessment compared to quantitative image analysis. BMC Cancer. 2009 May 29;9:165


Breast Analyte Control^{DR} **ER**





Breast Analyte Control^{DR} **PR**





QC Scoring: ER/PR

The heterogeneous cells are the key to determining how significantly the assay performance fluctuates. If the assay over stains or under stains, more or less of the cells in the heterogeneous cores will stain. For both ER and PR there are three cores with heterogeneous expression.

Due to the heterogeneous expression in the cores Quickscore or Allred scoring is not useful in assessing the cells. This is specifically for clinical assessment in tissue.

As the heterogeneous cell cores provide the utility for the control, a QC scoring method needs to be applied rather than a clinical one.

The table below has an example of a scoring method that easily conveys assay performance to the pathologist.

QC score	Result	Definition
1	Query/ fail	Control not staining as normal. Too weak. Repeat if necessary upon review of test sample.
2	Pass	Performance lower than expected but within tolerance
3	Pass	Performance as expected
4	Pass	Performance higher than expected but within tolerance
5	Query/ fail	Control not staining as normal. Too strong. Repeat if necessary upon review of test sample.



Also Available from HistoCyte Laboratories Ltd

Targets	Product Name	Format	Code
HPV/p16		Slide (2)	HCL001
	HPV/p16 Analyte Control ^{DR} (Four core with dynamic range of HPV gene copies)	Slide (5)	HCL002
		Block	HCL003
		Slide (2)	HCL004
	HPV/p16 Analyte Control (Three core with standard range of HPV gene copies)	Slide (5)	HCL005
	Standard range of the Serie copies)	Block	HCL006
ALK		Slide (2)	HCL007
	ALK-Lung Analyte Control (Two core positive and negative for the EML4-ALK translocation)	Slide (5)	HCL008
		Block	HCL009
	ALK-Lymphoma Analyte Control (Two core positive and negative for the NPM-ALK translocation)	Slide (2)	HCL010
		Slide (5)	HCL011
		Block	HCL012
Breast Analyte Control		Slide (2)	HCL013
	Breast Analyte Control (Two cores, one positive for Her2, ER and PR. The other negative)	Slide (5)	HCL014
		Block	HCL015
	Breast Analyte Control ^{DR} (Five cores with a	Slide (2)	HCL016
	dynamic range of expression of Her2, ER and PR.	Slide (5)	HCL017
	Including negative control)	Block	HCL018

Contact us for more information or a FREE SAMPLE*

Email: info@histocyte.com Telephone: +44 (0) 191 603 1007

HistoCyte Laboratories Ltd Herschel Annex Newcastle University Newcastle upon Tyne NE1 7RU United Kingdom <u>www.histocyte.com</u> Email: <u>info@histocyte.com</u> Telephone: +44 (0) 191 603 1007



Quality in Control